

Application Memo

Hydrochloric Acid in Acidic Cleaning Agent

Industry	Inorganic chemical industry
Instrument	Automatic potentiometric titrator
Measurement method	GB/T 622

1. Overview

The diluted sample is titrated with 1mol/L sodium hydroxide. The endpoint is the maximum inflection on the titration curve. The hydrochloric acid concentration is calculated from the titration volume of the sodium hydroxide solution.


2. Apparatus

Main unit	Automatic potentiometric titrator (preamplifier STD)
Electrode	pH glass electrode Ceramic reference electrode

3. Reagents

Titrant	1mol/L sodium hydroxide solution
---------	----------------------------------

4. Example



-Measurement results-

	Sample (mL)	Titer (mL)	Concentration (%)
1	2.0	1.4099	2.57
2	2.0	1.4332	2.62
3	2.0	1.4191	2.59
Average			2.59
SD			0.02
RSD(%)			1

Please feel free to contact us for any further information.

<Contact> Kyoto Electronics Manufacturing Co., Ltd.
Overseas Sales & Marketing Sect.